

CFR21
Certified

FlashLink® Reusable USB Datalogger with External Probe



Laboratory / R&D



Pharmaceutical



Refrigeration Vans



Ware Housing



Green House



Fruits & Vegetables

● Features :



- Large, Easy to read LCD Display
- Remaining Battery Indicator (Low Battery)
- NIST Traceable and CE Mark
- USB Connection
- Single Channel, External Probe for Temperature
- User Programmable : Manual or Automatic Start, Sample Interval, Alarm Limits, Data Tags

Note : CFR21 Software at additional charges



Model : 20902

Model : 20903

● Specifications :

Probe Measurement Range	-40°C to 85°C (-40°F to 185°F)
Probe Size: Penetration tip (20902)	0.14in Dia x 3.98in L (3.5mm Dia x 101mm L), 3.28 ft (1m) Teflon cable
Blunt tip (20903)	0.15in Dia x 0.98in L (3.8mm Dia x 25mm L), 3.28 ft (1m) Teflon cable
Temperature Resolution	0.1°C / °F
Accuracy	±0.5°C from -10°C to 25°C, (±1.0°F from 14°F to 77°F) ±1.0°C above 25°C, (±2.0°F above 77°F) ±1.0°C below -10°C, (±2.0°F below 14°F)
Data Capacity	3823 data points
Data Download Time	2 seconds typical for 3823 points
Programable Alarm Levels	High & Low levels over entire range
Battery	(CR2032) 3.6V Lithium / 18 months
Dimensions / Body	1.575" x 3.189" x 0.575" (40mm x 81mm x 13mm)
Casing	PC/ABS Plastic
Certification	CE Mark, NIST Traceable, RoHS Compliant
Display	LCD, Red LED Alarm
Waterproof Rating	IP64 with silicon cap on
System Requirements	Windows PC with current operating system, Adobe Reader
Programming	Flash PDF Program Manager software or USB Program Manager software
Download Requirements	Direct USB connection



HTA INSTRUMENTATION (P) LTD.,

An ISO 9001: 2015 Certified Company & NABL Accredited Calibration Laboratory as per ISO/IEC 17025:2017

Your One Stop For Instrumentation Supply, Automation & Calibration
73, Ramachandra Agrahara, Near T.R. Mills, Chamaraipet, Bangalore - 560018, INDIA.

Phone : 080-26749750, 26759253, 26740681 E-mail : sales@htaipl.com Website : www.htaipl.com

