



 <p>HTA Committed to the Customer since 1989</p>	<p><b>HTA INSTRUMENTATION (P) LTD.,</b> An ISO 9001: 2015 Certified Company &amp; NABL Accredited Calibration Laboratory as per ISO/IEC 17025:2017 <b>Your One Stop For Instrumentation Supply, Automation &amp; Calibration</b> # 73, Ramachandra Agrahara, Near T.R. Mills, Chamarajpet, Bangalore - 560018, INDIA. Phone : 080-26749750, 26759253, 26740681 E-mail : sales@htaipl.com Website : www.htaipl.com</p>	 <p><b>KVQA</b> eiaci ISO 9001:2015 CERTIFIED Certificate No. IQSC202010014</p>
---	---	--

## Specifications

EQ - RCT - 3220E

Model		RCT - 3220E Series Resistivity Controller		
Display Mode		LCD		
Auxiliary Sensor		0.0500 (cm <sup>-1</sup> )		
Sensor coefficient Range		(0.05 ± 0.00) cm <sup>-1</sup>		
Range		(0 ~ 18.25) MΩ.cm		
Resolution		0.01 MΩ.cm		
Accuracy		2.0%(FS)		
Temperature compensation		Double NTC		
Temperature compensation		Temp : (0~50) °C Relative Humidity : ≤ 85%RH		
Transmitter output		Isolated / transferable / reversible/ transmitter mode/meter mode		
Control Output	Output Contact	ON / OFF Double Contact Relay Output		
	Load Capacity	AC 220V/AC110V 3A Max DC 24V 3A (MAX)		
Power Supply		RCT - 3200E	RCT - 3210E	RCT - 3220E / RCT - 3220
		DC 24V ± 4V	AC 110V ±10%	AC 220V / 240V ±10%
Power Frequency Hz		---	50 /60	50 /60
Probe		Electrode Included (EQ-RCT3132-13)		
Meter Power Consumption		<2.5W		
Installation Method		Panel		
Dimension		48mm × 96mm × 80mm (H×W×D)		
Hole Dimension		44mm × 92mm (H×W)		
Weight		0.27kg		

Page - 2 / 2